

ACCURUN[®] 25 Series 2000

Multi-Marker ToRCH IgG Positive Control

For the DiaSorin LIAISON[®] Instrument

ACCURUN 25 Series 2000 ToRCH IgG Positive Control is formulated for use with *in vitro* diagnostic test kits for the qualitative determination of IgG antibodies to Toxoplasma gondii, Rubella virus, Cytomegalovirus (CMV) and Herpes Simplex Virus (HSV 1/2).

ACCURUN controls are specially formulated to exhibit weak reactivity in true patient-like matrices to pressure-test assay performance near critical clinical decision points. ACCURUN independent controls offer sensitive detection of subtle shifts in testing trends and mitigate the risk of reporting false results.



Low-Positive Controls

Specifically designed to be weakly reactive, low-positive controls, ACCURUN truly challenges an assay's performance at critical decision points.

Patient-Like Matrices

ACCURUN controls are formulated to mimic authentic patient samples, as encountered in a daily testing environment.

Third-Party/Independent

ACCURUN controls are designed to deliver an independent and unbiased assessment of performance.

Lot To Lot Consistency

Produced in large lot sizes under ISO 13485 and cGMP conditions, ACCURUN controls can be used for long-term QC monitoring.

Extended Shelf Life

ACCURUN serology controls are stable for at least 18 months from date of manufacture and at least 60 days once opened.

Ease of Use

Controls are offered in ready-to-use liquid format, enabling streamlined lab workflows.

Material No.	Kit Size	Matrix	Markers	Regulatory Status
2000-0074	3 x 1.0 mL	Human Serum or Plasma	IgG antibodies to Toxoplasma gondii, Rubella virus, Cytomegalovirus (CMV), Herpes Simplex Virus (HSV 1/2)	CE-IVD

Order: 800.377.9684 | +1 207.892.1300 | CDX-Sales@lgcgroup.com

Learn more by visiting www.seracare.com



CE 2797

MKT-00757 Rev. 2

Distributed by Abacus dx

1800 ABACUS (AUS) 0800 222 170 (NZ) | info@abacusdx.com | www.abacusdx.com

abacus dx