

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Tracer iodination date: 14 May 2026

Kit lot number: 3INK370

Kit expiry date: 10 July 2026

125I - Insulin	Lot No.	IT370R	Expiry Date:	10 JUL 2026
Assay Buffer	Lot No.	MIAB176	Expiry Date:	20 SEP 2027
Anti-human IgG	Lot No.	203PG	Expiry Date:	03 SEP 2027
Calibrators	Lot No.	IMS41A	Expiry Date:	01 DEC 2027
Negative Control	Lot No.	IN133	Expiry Date:	16 JUL 2026
Kit Controls	Lot No.	IMCa41A/IMCb41A	Expiry Date:	01 DEC 2027

Total cpm in 25 µl 125I - Insulin 28040

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.7		
INS-Ab calibrator:			
0.4 u/mL	2.3		
1 u/mL	4.3		
10 u/mL	15.2		
50 u/mL	55.9		
Control I	4.2	0.93	0.5 - 1.4
Control II	19.1	12.4	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 20 May 2026

Performed by: PP J DAVIES

Signature: A. ifrane

Position: Senior Scientist

Authorised by: DR G Flood

Signature: [Signature]

Position: Head of Quality Control

Date: 26TH MAY 2026