

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Tracer iodination date: 11 June 2026

Kit lot number: 3INK371

Kit expiry date: 07 August 2026

125I - Insulin	Lot No. IT371R	Expiry Date: 07 AUG 2026
Assay Buffer	Lot No. MIAB177	Expiry Date: 19 OCT 2027
Anti-human IgG	Lot No. 204PG	Expiry Date: 31 OCT 2027
Calibrators	Lot No. IMS41A	Expiry Date: 01 DEC 2027
Negative Control	Lot No. IN134	Expiry Date: 19 APR 2027
Kit Controls	Lot No. IMCa41A/IMCb41A	Expiry Date: 01 DEC 2027


Total cpm in 25 µl 125I - Insulin 29194

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.7		
INS-Ab calibrator:			
0.4 u/mL	2.3		
1 u/mL	4.0		
10 u/mL	16.1		
50 u/mL	58.1		
Control I	4.1	1.10	0.5 – 1.4
Control II	18.9	11.7	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.


Assay Date: 16 Jun 2026

Performed by: J DAVIES

Signature: 

Position: Senior Scientist

Authorised by: DR G Flood

Signature: 

Position: Head of Quality Control

Date: 16TH JUN 2026