

Title: **Quality Control Record Sheet
Fast IA-2 Ab ELISA Kit QC Data**

Document: **wi/qcrs/0.28b**
 Revision: **0**
 Date: **20 September 2022**
 Page **1 of 1**

Prepared by: **G FLOOD**
 Approved by: **L CARR**

Kit lot number: FKIE175B Kit expiry date: 05 January 2026

Kit manufacture date: 06 January 2025

Reagent Lot Numbers:

IA-2 Coated Wells: <u>IE237</u>	Expiry date: <u>16 MAR 2026</u>
IA-2-Biotin: <u>IEB166</u>	Expiry date: <u>15 APR 2026</u>
Recon. Buffer for IA-2-Biotin: <u>IBB186</u>	Expiry date: <u>18 MAR 2026</u>
Streptavidin Peroxidase: <u>GEP673</u>	Expiry date: <u>21 APR 2026</u>
Diluent for SA-POD: <u>PB182</u>	Expiry date: <u>27 JUN 2026</u>
Reaction Enhancer: <u>IES130</u>	Expiry date: <u>03 OCT 2026</u>
Peroxidase Substrate: <u>PTS185S</u>	Expiry date: <u>27 APR 2027</u>
Stop Solution: <u>HESP134</u>	Expiry date: <u>19 AUG 2026</u>
Concentrated Wash Solution: <u>CW410</u>	Expiry date: <u>20 FEB 2027</u>
Calibrators: <u>IC175b</u>	Expiry date: <u>05 JAN 2026</u>
Controls: <u>INC175b/IPC175b</u>	Expiry date: <u>05 JAN 2026</u>

Curve and Controls: Incubation temp: +21 °C

Calibrator Concentration (units per mL)	Absorbance at 450 nm	Absorbance at 405 nm
Negative control	0.088	0.025
7.5	0.149	0.042
35	0.394	0.112
120	1.693	0.488
350	>4	1.620
4000	>4	2.445

Control	Absorbance at 450 nm	Absorbance at 405 nm	Concentration at 450 nm (units per mL)	Concentration at 405 nm (units per mL)	Range (units per mL)
Positive I	1.735	0.499	121.8	121.7	125 ± 40

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay performed by: Fiona Ryrie
 Signature: *Fiona Ryrie*
 Position: Scientist
 Date: 29 October 2024

Authorised by: DR G FLOOD
 Signature: *G Flood*
 Position: Head of Quality Control
 Date: 6th JAN 2025