

Title: **Quality Control Record Sheet  
Insulin Ab 3 QC Data**

Prepared by: Geoff Flood  
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b  
Revision: 2  
Date: 09.01.14  
Page 1 of 1

Tracer iodination date: 31 October 2024

Kit lot number: 3INK350

Kit expiry date: 27 December 2024

125I - Insulin	Lot No.	IT350R	Expiry Date:	27 DEC 2024
Assay Buffer	Lot No.	MIAB157	Expiry Date:	10 FEB 2026
Anti-human IgG	Lot No.	194PG	Expiry Date:	17 APR 2026
Calibrators	Lot No.	IMS38E	Expiry Date:	26 FEB 2025
Negative Control	Lot No.	IN132	Expiry Date:	04 JUL 2025
Kit Controls	Lot No.	IMCa38E/IMCb38E	Expiry Date:	26 FEB 2025

Total cpm in 25 µl 125I - Insulin 29393

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.6		
INS-Ab calibrator:			
0.4 u/mL	1.8		
1 u/mL	3.6		
10 u/mL	15.0		
50 u/mL	57.8		
Control I	3.6	1.0	0.5 - 1.7
Control II	17.9	11.7	8 - 18

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 06 Nov 2024

Performed by: A HUCKLE

Signature: A. Huckle

Position: Senior Technician

Authorised by: DR G Flood

Signature: [Signature]

Position: Head of Quality Control

Date: 6<sup>th</sup> Nov 2024