

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Tracer iodination date: 16 May 2024

Kit lot number: 3INK344


Kit expiry date: 12 July 2024


125I - Insulin	Lot No.	IT344R	Expiry Date:	12 JUL 2024
Assay Buffer	Lot No.	MIAB152	Expiry Date:	25 AUG 2025
Anti-human IgG	Lot No.	189PG	Expiry Date:	27 SEP 2025
Calibrators	Lot No.	IMS38D	Expiry Date:	26 FEB 2025
Negative Control	Lot No.	IN132	Expiry Date:	04 JUL 2025
Kit Controls	Lot No.	IMCa38D/IMCb38D	Expiry Date:	26 FEB 2025

Total cpm in 25 µl 125I - Insulin 28321

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.5		
INS-Ab calibrator:			
0.4 u/mL	1.7		
1 u/mL	3.5		
10 u/mL	15.5		
50 u/mL	58.5		
Control I	3.7	1.3	0.5 - 1.7
Control II	18.7	11.9	8 - 18

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 22 May 2024
Performed by: J DAVIES
Signature: 
Position: Senior Scientist

Authorised by: DR G FLOOD
Signature: 
Position: Head of Quality Control
Date: 22ND MAY 2024