

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Tracer iodination date: 03 October 2024

Kit lot number: 3INK349

Kit expiry date: 29 November 2024

125I - Insulin	Lot No.	IT349R	Expiry Date:	29 NOV 2024
Assay Buffer	Lot No.	MIAB156	Expiry Date:	06 JAN 2026
Anti-human IgG	Lot No.	193PG	Expiry Date:	26 FEB 2026
Calibrators	Lot No.	IMS38E	Expiry Date:	26 FEB 2025
Negative Control	Lot No.	IN132	Expiry Date:	04 JUL 2025
Kit Controls	Lot No.	IMCa38E/IMCb38E	Expiry Date:	26 FEB 2025

Total cpm in 25 µl 125I - Insulin 28171

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.6		
INS-Ab calibrator:			
0.4 u/mL	1.8		
1 u/mL	3.8		
10 u/mL	15.5		
50 u/mL	60.8		
Control I	3.7	0.93	0.5 - 1.7
Control II	19.4	12.2	8 - 18

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 08 Oct 2024

Authorised by: DR G Flood

Performed by: J DAVIES

Signature: [Signature]

Signature: [Signature]

Position: Head of Quality Control

Position: Senior Scientist

Date: 8th Oct 2024