

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Tracer iodination date: 02 January 2025

Kit lot number: 3INK352


Kit expiry date: 28 February 2025


125I - Insulin	Lot No.	IT352R	Expiry Date:	28 FEB 2025
Assay Buffer	Lot No.	MIAB159	Expiry Date:	06 APR 2026
Anti-human IgG	Lot No.	194PG	Expiry Date:	17 APR 2026
Calibrators	Lot No.	IMS39A	Expiry Date:	27 FEB 2026
Negative Control	Lot No.	IN132	Expiry Date:	04 JUL 2025
Kit Controls	Lot No.	IMCa39A/IMCb39A	Expiry Date:	27 FEB 2026

Total cpm in 25 µl 125I - Insulin 30232

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.6		
INS-Ab calibrator:			
0.4 u/mL	2.4		
1 u/mL	3.5		
10 u/mL	16.8		
50 u/mL	59.1		
Control I	3.5	1.00	0.5 - 1.5
Control II	19.7	11.7	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 08 Jan 2025
Performed by: J DAVIES
Signature: 
Position: Senior Scientist

Authorised by: DR G FLOOD
Signature: 
Position: Head of Quality Control
Date: 8TH JAN 2025