

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Tracer iodination date: 16 April 2026

Kit lot number: 3INK369

Kit expiry date: 12 June 2026

125I - Insulin	Lot No. IT369R	Expiry Date: 12 JUN 2026
Assay Buffer	Lot No. MIAB175	Expiry Date: 25 AUG 2027
Anti-human IgG	Lot No. 202PG	Expiry Date: 20 JUL 2027
Calibrators	Lot No. IMS41A	Expiry Date: 01 DEC 2027
Negative Control	Lot No. IN133	Expiry Date: 16 JUL 2026
Kit Controls	Lot No. IMCa41A/IMCb41A	Expiry Date: 01 DEC 2027

Total cpm in 25 μ l ¹²⁵I - Insulin 29487

<u>Sample</u>	<u>% ¹²⁵I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	<u>0.7</u>		
INS-Ab calibrator:			
0.4 u/mL	2.2		
1 u/mL	3.9		
10 u/mL	14.9		
50 u/mL	55.3		
Control I	3.9	1.00	0.5 - 1.4
Control II	17.4	11.5	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 22 Apr 2026

Performed by: A Huckle

Signature: A. Huckle

Position: Senior Technician

Authorised by: DR G Flood

Signature: [Signature]

Position: Head of Quality Control

Date: 22nd APR 2026