

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: **wi/qcrs/0.08b**
Revision: **2**
Date: **09.01.14**
Page **1 of 1**

Prepared by: **Geoff Flood**
Approved by: **Janet Bradbury**

Tracer iodination date: 06 August 2026

Kit lot number: 3INK373

Kit expiry date: 02 October 2026

125I - Insulin	Lot No. IT373R	Expiry Date: 02 OCT 2026
Assay Buffer	Lot No. MIAB178	Expiry Date: 17 NOV 2027
Anti-human IgG	Lot No. 205PG	Expiry Date: 13 JAN 2028
Calibrators	Lot No. IMS41D	Expiry Date: 01 DEC 2027
Negative Control	Lot No. IN134	Expiry Date: 19 APR 2027
Kit Controls	Lot No. IMCa41D/IMCb41D	Expiry Date: 01 DEC 2027

Total cpm in 25 µl 125I - Insulin 27911

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	<u>0.7</u>		
INS-Ab calibrator:			
0.4 u/mL	2.2		
1 u/mL	4.1		
10 u/mL	16.0		
50 u/mL	57.5		
Control I	4.1	1.00	0.5 – 1.4
Control II	18.7	11.6	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 11 Aug 2026

Performed by: J DAVIES

Signature: 

Position: Senior Scientist

Authorised by: DR G flood

Signature: 

Position: Head of Quality Control

Date: 12th AUG 2026