

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Tracer iodination date: 03 September 2026

Kit lot number: 3INK374

Kit expiry date: 30 October 2026

125I - Insulin	Lot No. IT374R	Expiry Date: 30 OCT 2026
Assay Buffer	Lot No. MIAB179	Expiry Date: 14 DEC 2027
Anti-human IgG	Lot No. 205PG	Expiry Date: 13 JAN 2028
Calibrators	Lot No. IMS41D	Expiry Date: 01 DEC 2027
Negative Control	Lot No. IN134	Expiry Date: 19 APR 2027
Kit Controls	Lot No. IMCa41D/IMCb41D	Expiry Date: 01 DEC 2027

Total cpm in 25 µl 125I - Insulin 28886

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.6		
INS-Ab calibrator:			
0.4 u/mL	2.2		
1 u/mL	4.2		
10 u/mL	15.8		
50 u/mL	56.6		
Control I	4.1	0.94	0.5 - 1.4
Control II	19.0	12.0	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 08 Sep 2026

Performed by: J DAVIES

Signature: 

Position: Senior Scientist

Authorised by: DR G Flood

Signature: 

Position: Head of Quality Control

Date: 9TH SEP 2026