

Title: **Quality Control Record Sheet**
Insulin Ab 3 QC Data

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Tracer iodination date: 01 October 2026

Kit lot number: 3INK375

Kit expiry date: 27 November 2026

125I - Insulin	Lot No.	IT375R	Expiry Date:	27 NOV 2026
Assay Buffer	Lot No.	MIAB180	Expiry Date:	10 JAN 2028
Anti-human IgG	Lot No.	207PG	Expiry Date:	04 FEB 2028
Calibrators	Lot No.	IMS41D	Expiry Date:	01 DEC 2027
Negative Control	Lot No.	IN135	Expiry Date:	30 SEP 2027
Kit Controls	Lot No.	IMCa41D/IMCb41D	Expiry Date:	01 DEC 2027

Total cpm in 25 µl 125I - Insulin 28949

<u>Sample</u>	<u>% 125I - Insulin</u> <u>bound</u>	<u>u/mL</u>	<u>Range</u> <u>u/mL</u>
Negative control	1.1		
INS-Ab calibrator:			
0.4 u/mL	2.4		
1 u/mL	4.4		
10 u/mL	18.2		
50 u/mL	59.7		
Control I	4.3	0.95	0.5 - 1.4
Control II	21.7	12.1	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 07 Oct 2026

Performed by: JD J DAVIES

Signature: [Signature]

Position: Senior Scientist

Authorised by: DR G flood

Signature: [Signature]

Position: Head of Quality Control

Date: 12th Oct 2026