

Title: **Quality Control Record Sheet  
Insulin Ab 3 QC Data**

Document: wi/qcrs/0.08b

Revision: 2

Prepared by: Geoff Flood

Date: 09.01.14

Approved by: Janet Bradbury

Page 1 of 1

Tracer iodination date: 19 February 2026<sup>b</sup>

Kit lot number: 3INK367

Kit expiry date: 17 April 2026

125I - Insulin	Lot No. IT367R	Expiry Date: 17 APR 2026
Assay Buffer	Lot No. MIAB172	Expiry Date: 07 JUN 2027
Anti-human IgG	Lot No. 201PG	Expiry Date: 21 APR 2027
Calibrators	Lot No. IMS41A	Expiry Date: 01 DEC 2027
Negative Control	Lot No. IN133	Expiry Date: 16 JUL 2026
Kit Controls	Lot No. IMCa41A/IMCb41A	Expiry Date: 01 DEC 2027

Total cpm in 25 µl 125I - Insulin 29081

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.6		
INS-Ab calibrator:			
0.4 u/mL	2.0		
1 u/mL	3.9		
10 u/mL	16.3		
50 u/mL	53.4		
Control I	4.1	1.14	0.5 - 1.4
Control II	17.8	11.0	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 24 Feb 2026

Authorised by: DR G Flood

Performed by: J DAVIES

Signature: [Signature]

Signature: [Signature]

Position: Head of Quality Control

Position: Senior Scientist

Date: 24<sup>th</sup> FEB 2026