

Title: **Quality Control Record Sheet**
TRAb Coated Tubes QC Data

Document: wi/qcrs/0.23

Revision: 5

Prepared by: G Flood

Date: 11.05.21

Approved by: W Dowler

Page 1 of 1

Kit Lot Number: KTCT425A

Kit expiry date: 27 May 2026

125I –TSH Tracer:	Lot No. TCT389	Expiry Date 27 May 2026
TSH Receptor Coated Tubes:	Lot No. 2RC1101	Expiry Date 19 Jul 2026
Calibrators:	Lot No. 2CTC300	Expiry Date 26 Jun 2027
Negative Control:	Lot No. 2CTC300N	Expiry Date 26 Jun 2027
Positive Control:	Lot No. 2CTC300P	Expiry Date 26 Jun 2027
Concentrated Wash Solution:	Lot No. TEW148	Expiry Date 11 Aug 2028
Start Buffer:	Lot No. 2aAD113	Expiry Date 17 Jun 2027

Total counts: 84367 cpm

Temp. 21 °C


	<u>IU/L</u>	<u>% bound</u>	<u>% inhibition</u>	
Calibrator				
1	1	13.59	13	
2	2	12.15	22	
3	8	7.20	54	
4	40	2.74	82	
CONTROL	<u>% bound</u>	<u>% inhibition</u>	<u>IU/L</u>	<u>Range</u>
I (Negative)	15.66	-	-	
II (Positive)	10.59	32	3.3	3.9 +/- 1.2 IU/L

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 25 Mar 2026

Authorised by: SARAH NUTE

Performed by: Wendy Dowler

Signature: 

Signature: 

Position: pp Head of Quality Control

Position: **Section Leader**

Date: 25 MAR 2026