

Title: **Quality Control Record Sheet**
TRAb Coated Tubes QC Data

Document: wi/qcrs/0:23
Revision: 5
Date: 11.05.21
Page 1 of 1

Prepared by: G Flood
Approved by: W Dowler


Kit Lot Number: KTCT426 Kit expiry date: 24 Jun 2026


125I -TSH Tracer:	Lot No. TCT390	Expiry Date 24 Jun 2026
TSH Receptor Coated Tubes:	Lot No. 2RC1101	Expiry Date 19 Jul 2026
Calibrators:	Lot No. 2CTC301	Expiry Date 28 Oct 2027
Negative Control:	Lot No. 2CTC301N	Expiry Date 28 Oct 2027
Positive Control:	Lot No. 2CTC301P	Expiry Date 28 Oct 2027
Concentrated Wash Solution:	Lot No. TEW148	Expiry Date 11 Aug 2028
Start Buffer:	Lot No. 2aAD113	Expiry Date 17 Jun 2027

Total counts: 84186 cpm Temp. 21 °C

	IU/L	% bound	% inhibition	
Calibrator				
1	1	13.07	14	
2	2	11.41	25	
3	8	6.47	58	
4	40	2.51	84	
CONTROL	% bound	% inhibition	IU/L	Range
I (Negative)	15.28	-	-	
II (Positive)	9.55	38	3.5	3.5 +/- 1.2 IU/L

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 23 Apr 2026
Performed by: Wendy Dowler
Signature: 
Position: **Section Leader**

Authorised by: SARAH NUTE
Signature: 
Position: pp Head of Quality Control
Date: 23 APR 2026