

Title: **Quality Control Record Sheet**
TRAb Coated Tubes QC Data

Prepared by: G Flood
Approved by: W Dowler

Document: wi/qcrs/0.23
Revision: 5
Date: 11.05.21
Page 1 of 1

Kit Lot Number: **KTCT430**

Kit expiry date: **14 Oct 2026**

¹²⁵ I –TSH Tracer:	Lot No. TCT394	Expiry Date 14 Oct 2026
TSH Receptor Coated Tubes:	Lot No. 2RC1109	Expiry Date 11 Nov 2026
Calibrators:	Lot No. 2CTC305	Expiry Date 28 Oct 2027
Negative Control:	Lot No. 2CTC305N	Expiry Date 28 Oct 2027
Positive Control:	Lot No. 2CTC305P	Expiry Date 28 Oct 2027
Concentrated Wash Solution:	Lot No. TEW150	Expiry Date 06 May 2029
Start Buffer:	Lot No. 2aAD113	Expiry Date 17 Jun 2027

Total counts: **84054** cpm

Temp. **21** °C

	<u>IU/L</u>	<u>% bound</u>	<u>% inhibition</u>	
Calibrator				
1	1	12.72	13	
2	2	11.40	22	
3	8	5.72	61	
4	40	2.20	85	
CONTROL	<u>% bound</u>	<u>% inhibition</u>	<u>IU/L</u>	<u>Range</u>
I (Negative)	14.66	-	-	
II (Positive)	9.21	37	3.6	3.5 +/- 1.2 IU/L

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.


Assay Date: **12 Aug 2026**

Performed by: **Wendy Dowler**

Signature: 

Position: **Section Leader**

Authorised by: 

Signature: 

Position: **Head of Quality Control**

Date: **13th AUG 2026**