

Kit Lot Number:
KTCT431

Kit expiry date:
11 November 2026

125I –TSH Tracer:	Lot No.	TCT395	Expiry Date	11 November 2026
TSH Receptor Coated Tubes:	Lot No.	2RC1111	Expiry Date	07 December 2026
Calibrators:	Lot No.	2CTC306	Expiry Date	28 October 2027
Negative Control:	Lot No.	2CTC306N	Expiry Date	28 October 2027
Positive Control:	Lot No.	2CTC306P	Expiry Date	28 October 2027
Concentrated Wash Solution:	Lot No.	TEW150	Expiry Date	06 May 2029
Start Buffer:	Lot No.	2aAD113	Expiry Date	17 June 2027

Total counts:
86620
cpm

Temp.
21
°C

	IU/L	% bound	% inhibition	
Calibrator				
1	1	14.28	13	
2	2	12.39	24	
3	8	6.93	58	
4	40	2.45	85	
CONTROL	% bound	% inhibition	IU/L	Range
I (Negative)	16.38	-	-	
II (Positive)	10.44	36	3.4	3.5 +/- 1.2 IU/L

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.


Assay Date:
09 September 2026

Performed by:
Wendy Dowler

Signature:


Position:
Section Leader

Authorised by:
DR G Flood

Signature:


Position:
Head of Quality Control

Date:
10th SEP 2026