Title: **Quality Control Record Sheet**
TRAb Coated Tubes QC Data

Prepared by: G Flood
Approved by: W Dowler

Document: wi/qcrs/0.23
Revision: 5
Date: 11.05.21
Page 1 of 1

Kit Lot Number: **KTCT432A** Kit expiry date: **09 December 2026**

^{125}I –TSH Tracer:	Lot No.	**TCT396**	Expiry Date	**09 December 2026**
TSH Receptor Coated Tubes:	Lot No.	**2RC1115**	Expiry Date	**19 January 2027**
Calibrators:	Lot No.	**2CTC307**	Expiry Date	**28 October 2027**
Negative Control:	Lot No.	**2CTC307N**	Expiry Date	**28 October 2027**
Positive Control:	Lot No.	**2CTC307P**	Expiry Date	**28 October 2027**
Concentrated Wash Solution:	Lot No.	**TEW150**	Expiry Date	**06 May 2029**
Start Buffer:	Lot No.	**2aAD113**	Expiry Date	**17 June 2027**

Total counts: **84358** cpm Temp. **21** °C

	IU/L	**% bound**	**% inhibition**	
Calibrator				
1	**1**	12.66	16	
2	**2**	11.37	24	
3	**8**	5.97	60	
4	**40**	2.25	85	
CONTROL	**% bound**	**% inhibition**	**IU/L**	**Range**
I (Negative)	15.06	-	-	
II (Positive)	9.18	39	3.7	**3.5 +/- 1.2 IU/L**

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: **07 October 2026**
Performed by: **Morwen Williams**
Signature: MWilliams
Position: **Scientist**

Authorised by: SARAH NUTE
Signature: SRNute
Position: pp **Head of Quality Control**
Date: 09 OCT 2026.