

Title: **Quality Control Record Sheet**
TRAb Coated Tubes QC Data

Document: wi/qcrs/0.23
Revision: 5
Date: 11.05.21
Page 1 of 1

Prepared by: G Flood
Approved by: W Dowler

Kit Lot Number: KTCT424

Kit expiry date: 29 Apr 2026

125I -TSH Tracer:	Lot No. TCT388	Expiry Date 29 Apr 2026
TSH Receptor Coated Tubes:	Lot No. 2RC1098	Expiry Date 24 May 2026
Calibrators:	Lot No. 2CTC299	Expiry Date 26 Jun 2027
Negative Control:	Lot No. 2CTC299N	Expiry Date 26 Jun 2027
Positive Control:	Lot No. 2CTC299P	Expiry Date 26 Jun 2027
Concentrated Wash Solution:	Lot No. TEW148	Expiry Date 11 Aug 2028
Start Buffer:	Lot No. 2aAD112	Expiry Date 02 Aug 2026

Total counts: 81905 cpm

Temp. 21 °C

	<u>IU/L</u>	<u>% bound</u>	<u>% inhibition</u>	
Calibrator				
1	1	11.99	14	
2	2	10.83	22	
3	8	6.52	53	
4	40	2.61	81	
CONTROL	<u>% bound</u>	<u>% inhibition</u>	<u>IU/L</u>	<u>Range</u>
I (Negative)	13.88	-	-	
II (Positive)	8.93	36	3.9	3.9 +/- 1.2 IU/L

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 26 Feb 2026

Performed by: Claire Fisher

Signature: pp [Signature]

Position: Laboratory Technician

Authorised by: SARAH NUTE

Signature: [Signature]

Position: pp Head of Quality Control

Date: 27 FEB 2026.