

Title: **Quality Control Record Sheet**
RBA AChRAb Assay QC data


Document: **wi/qcrs/0.14a**
Revision: **7**
Date: **6 Sep 2018**
Page: **2 of 2**



Prepared by: Geoff Flood
Approved by: Rachael Price

Kit lot no: **KRBA349** Expiry date: **09 Jul 2024**

Assay date (weeks after receptor labelling)	Decay factor (A)
Up to 1 week	1.0
+ 1 – 2 weeks	1.1
+ 2 – 3 weeks	1.2
+ 3 – 4 weeks	1.3
+ 4 – 5 weeks	1.4
+ 5 – 6 weeks	1.5
+ 6 – 7 weeks	1.6
+ 7 – 8 weeks	1.75
+ 8 – 9 weeks	1.9

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay date: **10 May 2024**
Performed by: **A Lee**
Signature: 
Position: **Principal Technician**

Authorised by: 
Signature: 
Position: **Head of Quality Control**
Date: **13TH MAY 2024**