

Re: AChRAb Standard Curve (lot no. ASC20ae; expiry date. 12 Feb 2025)

To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.

No ¹²⁵I decay correction is needed.


Our QC data for the standard curve, obtained with AChR lot **RBA349** are as follows:

Standard	Actual value at QC
1	0.26 nmol/litre
2	0.92 nmol/litre
3	3.4 nmol/litre
4	9.1 nmol/litre
Value of positive control 118Pe read off standard curve = 4.3 nmol/litre	
(Range: 2.5-5.9 nmol/litre)	


Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.


Assay Date: 10 May 2024

Performed by: A Lee

Signature: 

Position: Principal Technician

Authorised by: 

Signature: 

Position: Head of Quality Control

Date: 13TH MAY 2024