

Title: **Quality Control Record Sheet
AChRAb Standard Curve QC Data**

Prepared by: Geoff Flood
Approved by: Rachael Price

Document: **wi/qcrs/0.14SCb**
Revision: **3**
Date: **6 Sep 2018**
Page **1 of 1**

Re: AChRAb Standard Curve (lot no. ASC20ag; expiry date. 07 Apr 2025)

To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.

No ¹²⁵I decay correction is needed.


Our QC data for the standard curve, obtained with AChR lot **RBA351** are as follows:

Standard	Actual value at QC
1	0.37 nmol/litre
2	1.2 nmol/litre
3	3.7 nmol/litre
4	8.4 nmol/litre
Value of positive control 118Pg read off standard curve = 4.3 nmol/litre	
(Range: 2.5-5.9 nmol/litre)	


Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.


Assay Date: 04 Jul 2024

Performed by: A Lee

Signature: 

Position: Principal Technician

Authorised by: 

Signature: 

Position: Head of Quality Control

Date: 5TH JULY 2024