

Title: **Quality Control Record Sheet
AChRAb Standard Curve QC Data**

Document: **wi/qcrs/0.14SCb**

Revision: **3**

Prepared by: Geoff Flood

Date: **6 Sep 2018**

Approved by: Rachael Price

Page **1 of 1**

Re: AChRAb Standard Curve (lot no. ASC21e; expiry date. 05 Oct 2025)

To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.

No ¹²⁵I decay correction is needed.

Our QC data for the standard curve, obtained with AChR lot **RBA357** are as follows:

Standard	Actual value at QC	
1	0.34	nmol/litre
2	1.0	nmol/litre
3	3.5	nmol/litre
4	8.2	nmol/litre
Value of positive control 118Pr read off standard curve = 4.3 nmol/litre		
(Range: 2.5-5.9 nmol/litre)		

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 19 Dec 2024

Authorised by: DR G Flood

Performed by: A Lee

Signature: [Signature]

Signature: [Signature]

Position: Head of Quality Control

Position: Principal Technician

Date: 20th DEC 2024