

Title: **Quality Control Record Sheet
AChRab Standard Curve QC Data**

Document: **wi/qcrs/0.14SCb**
 Revision: **3**
 Date: **6 Sep 2018**
 Page **1 of 1**

Prepared by: Geoff Flood
 Approved by: Rachael Price

Re: AChRab Standard Curve (lot no. ASC21f; expiry date. 19 Oct 2025)

To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.

No ¹²⁵I decay correction is needed.

Our QC data for the standard curve, obtained with AChR lot **RBA358** are as follows:

Standard	Actual value at QC
1	0.25 nmol/litre
2	1.0 nmol/litre
3	4.0 nmol/litre
4	8.9 nmol/litre
Value of positive control 118Pt read off standard curve = 4.9 nmol/litre	
(Range: 2.5-5.9 nmol/litre)	

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 16 Jan 2025
 Performed by: A Lee
 Signature: *M. Lee*
 Position: Principal Technician

Authorised by: *DR G Flood*
 Signature: *[Signature]*
 Position: Head of Quality Control
 Date: 17th JAN 2025