

Title: **Quality Control Record Sheet
AChRAb Standard Curve QC Data**

Document: **wi/qcrs/0.14SCb**

Revision: **3**

Prepared by: Geoff Flood

Date: **6 Sep 2018**

Approved by: Rachael Price

Page **1 of 1**

Re: AChRAb Standard Curve (lot no. ASC21h; expiry date. 16 Dec 2025)

To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.

No ¹²⁵I decay correction is needed.


Our QC data for the standard curve, obtained with AChR lot **RBA360** are as follows:

Standard	Actual value at QC	
1	0.22	nmol/litre
2	0.95	nmol/litre
3	3.4	nmol/litre
4	8.3	nmol/litre
Value of positive control 118Pv read off standard curve = 4.6 nmol/litre		
(Range: 2.5-5.9 nmol/litre)		

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 13 Mar 2025

Performed by: A Lee

Signature: 

Position: Principal Technician

Authorised by: 

Signature: 

Position: Head of Quality Control

Date: 14th MAR 2025