

Title:

Quality Control Record Sheet
Insulin Ab 3 QC Data

Prepared by: Geoff Flood

Approved by: Janet Bradbury

Document: wi/qcrs/0.08b

Revision: 2

Date: 09.01.14

Page 1 of 1


Tracer iodination date: 05 January 2026Kit lot number: 3INK365Kit expiry date: 03 March 2026

125I - Insulin	Lot No.	IT365R	Expiry Date:	03 MAR 2026
Assay Buffer	Lot No.	MIAB170	Expiry Date:	09 APR 2027
Anti-human IgG	Lot No.	201PG	Expiry Date:	21 APR 2027
Calibrators	Lot No.	IMS40E	Expiry Date:	26 FEB 2027
Negative Control	Lot No.	IN133	Expiry Date:	16 JUL 2026
Kit Controls	Lot No.	IMCa40E/IMCb40E	Expiry Date:	26 FEB 2027

Total cpm in 25 μ L 125I - Insulin 29940

Sample	% 125I - Insulin bound	u/mL	Range u/mL
Negative control	0.7		
INS-Ab calibrator:			
0.4 u/mL	2.6		
1 u/mL	5.0		
10 u/mL	18.4		
50 u/mL	58.4		
Control I	4.3	0.75	0.45 – 1.45
Control II	20.5	11.3	6 – 18

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 09 Jan 2026Authorised by: DR G FloodPerformed by: A HuckleSignature: Signature: A. HucklePosition: Head of Quality ControlPosition: Snr TechnicianDate: 12th JAN 2026