

Title: **Quality Control Record Sheet**
TRAb Coated Tubes QC Data

Document: wi/qcrs/0.23

Revision: 5

Prepared by: G Flood

Date: 11.05.21

Approved by: W Dowler

Page 1 of 1

Kit Lot Number: KTCT410A

Kit expiry date: 02 Apr 2025

125I –TSH Tracer:	Lot No.	TCT374	Expiry Date	02 Apr 2025
TSH Receptor Coated Tubes:	Lot No.	2RC1068	Expiry Date	21 Apr 2025
Calibrators:	Lot No.	2CTC285	Expiry Date	02 Dec 2026
Negative Control:	Lot No.	2CTC285N	Expiry Date	02 Dec 2026
Positive Control:	Lot No.	2CTC285P	Expiry Date	02 Dec 2026
Concentrated Wash Solution:	Lot No.	TEW144	Expiry Date	22 Apr 2027
Start Buffer:	Lot No.	2aAD111	Expiry Date	27 Nov 2025

Total counts: 87541 cpm

Temp. 21 °C

	<u>IU/L</u>	<u>% bound</u>	<u>% inhibition</u>	
Calibrator				
1	1	13.62	15	
2	2	12.18	24	
3	8	6.08	62	
4	40	1.73	89	
CONTROL	<u>% bound</u>	<u>% inhibition</u>	<u>IU/L</u>	<u>Range</u>
I (Negative)	15.94	-	-	
II (Positive)	9.38	41	4.0	3.9 +/- 1.2 IU/L

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 29 Jan 2025

Authorised by: SARAH NUTE

Performed by: Hannah Sanyalou

Signature: [Signature]

Signature: [Signature]

Position: pp Head of Quality Control

Position: **Senior Scientist**

Date: 30 JAN 2025