

Title: **Quality Control Record Sheet**
TRAb Coated Tubes QC Data

Prepared by: G Flood
Approved by: W Dowler

Document: wi/qcrs/0.23
Revision: 5
Date: 11.05.21
Page 1 of 1

Kit Lot Number: KTCT428A

Kit expiry date: 19 Aug 2026

125I –TSH Tracer:	Lot No.	TCT392	Expiry Date	19 Aug 2026
TSH Receptor Coated Tubes:	Lot No.	2RC1107	Expiry Date	13 Oct 2026
Calibrators:	Lot No.	2CTC303	Expiry Date	28 Oct 2027
Negative Control:	Lot No.	2CTC303N	Expiry Date	28 Oct 2027
Positive Control:	Lot No.	2CTC303P	Expiry Date	28 Oct 2027
Concentrated Wash Solution:	Lot No.	TEW148	Expiry Date	11 Aug 2028
Start Buffer:	Lot No.	2aAD113	Expiry Date	17 Jun 2027

Total counts: 86828 cpm


Temp. 21 °C

	<u>IU/L</u>	<u>% bound</u>	<u>% inhibition</u>	
<u>Calibrator</u>				
1	1	13.91	15	
2	2	12.57	23	
3	8	7.13	56	
4	40	2.87	82	
<u>CONTROL</u>	<u>% bound</u>	<u>% inhibition</u>	<u>IU/L</u>	<u>Range</u>
I (Negative)	16.38	-	-	
II (Positive)	10.71	35	3.4	3.5 +/- 1.2 IU/L

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.


Assay Date: 17 Jun 2026

Performed by: Morwen Williams

Signature: 

Position: Scientist

Authorised by: SARAH NUTE

Signature: 

Position: pp Head of Quality Control

Date: 18 JUN 2026