

Title: **Quality Control Record Sheet**
TRAb Coated Tubes QC Data

Prepared by: G Flood
Approved by: W Dowler

Document: wi/qcrs/0.23
Revision: 5
Date: 11.05.21
Page 1 of 1

Kit Lot Number: KTCT422A

Kit expiry date: 11 Mar 2026

125I -TSH Tracer:	Lot No.	TCT386	Expiry Date	11 Mar 2026
TSH Receptor Coated Tubes:	Lot No.	2RC1096	Expiry Date	19 Apr 2026
Calibrators:	Lot No.	2CTC297	Expiry Date	26 Jun 2027
Negative Control:	Lot No.	2CTC297N	Expiry Date	26 Jun 2027
Positive Control:	Lot No.	2CTC297P	Expiry Date	26 Jun 2027
Concentrated Wash Solution:	Lot No.	TEW148	Expiry Date	11 Aug 2028
Start Buffer:	Lot No.	2aAD112	Expiry Date	02 Aug 2026

Total counts: 80748 cpm

Temp. 21 °C

	<u>IU/L</u>	<u>% bound</u>	<u>% inhibition</u>	
Calibrator				
1	1	12.30	15	
2	2	11.25	22	
3	8	6.76	53	
4	40	2.71	81	
CONTROL	<u>% bound</u>	<u>% inhibition</u>	<u>IU/L</u>	<u>Range</u>
I (Negative)	14.48	-	-	
II (Positive)	9.70	33	3.5	3.9 +/- 1.2 IU/L

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 07 Jan 2026
Performed by: Hannah Sanyaolu
Signature: H.Sanyaolu
Position: Senior Scientist

Authorised by: SARAH NUTE
Signature: SRNute
Position: pp Head of Quality Control
Date: 09 JAN 2026