

Title: **Quality Control Record Sheet**
TRAb Coated Tubes QC Data

Prepared by: G Flood
Approved by: W Dowler

Document: wi/qcrs/0.23
Revision: 5
Date: 11.05.21
Page: 1 of 1

Kit Lot Number: KTCT427

Kit expiry date: 22 Jul 2026

¹²⁵ I –TSH Tracer:	Lot No. TCT391	Expiry Date 22 Jul 2026
TSH Receptor Coated Tubes:	Lot No. 2RC1104	Expiry Date 01 Sep 2026
Calibrators:	Lot No. 2CTC302	Expiry Date 28 Oct 2027
Negative Control:	Lot No. 2CTC302N	Expiry Date 28 Oct 2027
Positive Control:	Lot No. 2CTC302P	Expiry Date 28 Oct 2027
Concentrated Wash Solution:	Lot No. TEW148	Expiry Date 11 Aug 2028
Start Buffer:	Lot No. 2aAD113	Expiry Date 17 Jun 2027

Total counts: 85797 cpm

Temp. 21 °C

	<u>IU/L</u>	<u>% bound</u>	<u>% inhibition</u>	
Calibrator				
1	1	12.90	14	
2	2	11.55	23	
3	8	6.98	54	
4	40	2.51	83	
CONTROL	<u>% bound</u>	<u>% inhibition</u>	<u>IU/L</u>	<u>Range</u>
I (Negative)	15.05	-	-	
II (Positive)	9.78	35	3.6	3.5 +/- 1.2 IU/L

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 20 May 2026

Performed by: Morwen Williams

Signature: 

Position: **Scientist**

Authorised by: SARAH NUTE

Signature: 

Position: pp **Head of Quality Control**

Date: 21 MAY 2026