

Title: **Quality Control Record Sheet
TRAb Coated Tubes QC Data**

Document: wi/qcrs/0.23
Revision: 5
Date: 11.05.21
Page: 1 of 1

Prepared by: G Flood
Approved by: W Dowler

Kit Lot Number: KTCT423

Kit expiry date: 01 Apr 2026

125I -TSH Tracer:	Lot No. TCT387	Expiry Date	01 Apr 2026
TSH Receptor Coated Tubes:	Lot No. 2RC1096A	Expiry Date	19 Apr 2026
Calibrators:	Lot No. 2CTC298	Expiry Date	26 Jun 2027
Negative Control:	Lot No. 2CTC298N	Expiry Date	26 Jun 2027
Positive Control:	Lot No. 2CTC298P	Expiry Date	26 Jun 2027
Concentrated Wash Solution:	Lot No. TEW148	Expiry Date	11 Aug 2028
Start Buffer:	Lot No. 2aAD112	Expiry Date	02 Aug 2026

Total counts: 85688 cpm

Temp. 22 °C

	<u>IU/L</u>	<u>% bound</u>	<u>% inhibition</u>	
Calibrator				
1	1	12.63	15	
2	2	11.24	24	
3	8	6.33	57	
4	40	2.23	85	
CONTROL	<u>% bound</u>	<u>% inhibition</u>	<u>IU/L</u>	<u>Range</u>
I (Negative)	14.85	-	-	
II (Positive)	8.99	39	4.0	3.9 +/- 1.2 IU/L

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 28 Jan 2026

Performed by: Wendy Dowler

Signature: 

Position: **Section Leader**

Authorised by: SARAH NUTE

Signature: 

Position: pp **Head of Quality Control**

Date: 29 JAN 2026